Jeruzalémská 1283/9, 110 52 Praha 1, Czech Republic Notified Body 1024



## EU TYPE-EXAMINATION CERTIFICATE No. 1024/E-021/2020

This EU type-examination certificate is issued to:

תנהנהנהנה ההנהנהנה וויות

Manufacturer:	Özbay Is Güvenligi Malzemeleri, San. Ve Tic. Ltd. Sti., Tersane Cad.No: 113-B, Karaköy/ Istanbul / Turkey
Identification number:	792
PPE description:	Filtering half mask to protect against particles
PPE product:	VENTO
Туре:	201 FFP2 V NR

It is certified that the manufacturer's technical file and above mentioned PPE product have been assessed and found to be in accordance with the essential health and safety requirements of Regulation (EU) 2016/425 of the European parliament and of the Council on personal protective equipment and repealing Council Directive 89/686/EEC, as recorded in Report No.1024/ZZ-021/2020, which is an integral part of this Certificate.

When examined the model was found to meet all of the relevant requirements of the appropriate harmonized standard(s):

EN 149:2001+A1:2009 Respiratory protective devices. Filtering half masks to protect against particles. Requirements, testing, marking (idt. ČSN EN 149:2002+A1:2009, ČSN EN 149+A1 OPRAVA 1:2018)

The certification was performed according to the certification scheme of Regulation (EU) 2016/425 Module B.

Marking and instructions have been assessed in the English language only.


Certificate is valid until 25. 9. 2025

For and on behalf of Occupational Safety Research Institute

Date of Issue: 25. 9. 2020



This certificate was issued in Czech and English versions. Both versions have the same validity.



Graphical appearance of CE mark. In the event of the involvement of a notified body in the stage of manufacturing check (category III, Module C2 or D) its distinguishing number shall be added to CE mark.

FFP2 NR 001 :2009 1024

EU TYPE-EXAMINATION CERTIFICATE No. 1024/E-021/2020

Description and picture of the PPE product:

half mask VENTO 201 FFP2 ۷ NR filtering provides protection of respiratory system of user against solid and liquid aerosols in accordance with the information supplied by the manufacturer.

The product meets the class FFP2 requirements.

The product is classified in category III and shall only be used in conjunction with conformity assessment procedure Module C2 within the meaning of Regulation (EU) 2016/425.